

**CERTIFICATE  
OF ACCREDITATION**



**Standards Council of Canada  
Conseil canadien des normes**

**CERTIFICAT  
D'ACCREDITATION**

**TÜV SÜD AMERICA INC.**  
10 Centennial Drive, Peabody, Massachusetts, 01960, USA

having been assessed and found to conform with the requirements of ISO/IEC 17020:2012, the Additional Requirements for Accreditation of Inspection Bodies (CAN-P-1608), and other requirements, and the conditions for accreditation established by SCC is hereby recognized as an


ayant été soumis à une évaluation et ayant été jugé conforme aux exigences d'ISO/CEI 17020:2012, aux *Exigences supplémentaires relatives à l'accréditation des organismes d'inspection* (CAN-P-1608) et autres exigences, ainsi qu'aux conditions d'accréditation établies par le CCN, est de fait reconnu comme étant un

**ACCREDITED INSPECTION BODY**

In specific subject areas and for key locations as identified in the scope of accreditation approved by SCC and found on the SCC website at [www.scc.ca](http://www.scc.ca).

**ORGANISME D'INSPECTION ACCREDITÉ**

dans les domaines d'activité et pour les établissements clés indiqués dans la portée d'accréditation approuvée par le CCN et figurant dans le site web du CCN au [www.ccn.ca](http://www.ccn.ca).



Accreditation number: / Numéro d'accréditation: 06009

Initial accreditation date: / Date de la première accréditation: 2005-08-05

Re-accreditation date: / Date de la réaccréditation: 2015-02-18

Expiry date: / Date d'expiration: 2018-08-05

*This certificate is valid until the date of expiration unless suspended, withdrawn or superseded by SCC. / Le présent certificat est valide jusqu'à la date d'expiration, à moins qu'il ne soit suspendu, retiré ou remplacé par le CCN.*

*Chantal Auay*

Vice-President – Accreditation Services / Vice-présidente – Services d'accréditation

To verify the validity of this certificate, please see the Directory of Accredited customers on [www.scc.ca](http://www.scc.ca).

Pour vérifier la validité du certificat, veuillez consulter le Répertoire des clients accrédités au [www.ccn.ca](http://www.ccn.ca).



## INSPECTION BODY ACCREDITATION PROGRAM (IBAP)

### Scope of Accreditation

**Accredited Legal Entity:** TÜV SÜD AMERICA INC.

Contact Name: Barry Quinlan

Address: 10 Centennial Drive  
Peabody, MA 01960  
U.S.A

Telephone: +1 978 573 2500

Fax: +1978 977 0182

Website: [www.tuv-sud-america.com/us-en](http://www.tuv-sud-america.com/us-en)

Email: bquinlan@tuvam.com

<b>Accreditation Number:</b>	06009
<b>Initial Accreditation Date:</b>	2005-08-05
<b>Date of Most Recent Reaccreditation:</b>	2015-02-18
<b>Accreditation Expiry Date:</b>	2018-08-05
<b>Accreditation Standards:</b>	ISO/IEC 17020:2012 SCC Requirements and Guidance, Inspection Body Accreditation Program
<b>Accreditation Sub-programs:</b>	SPE-1000 Field Evaluation/Special Inspection SPE-3000 Field Evaluation/Special Inspection

### Additional Fixed Office Locations:

Location	Country	Address	City
A	Canada	11 Gordon Collins Drive, Gormley, Ontario L0H 1G0, Canada	Gormley


### SPE-1000 Field Evaluation/Special Inspection



**Inspection Label:**



Product inspected to CSA SPE-1000 prior to April 01, 2010 was labeled in accordance with CSA-SPE-1000-99 as represented by the labels below.



**Scope of Accreditation:**

The performing of field evaluations/special inspections and the application of approval labels on electrical equipment and systems in accordance with all of the following:

As identified in the scope of CSA SPE-1000 Model Code for the field evaluation/special inspection of electrical equipment, but excluding:

- a) wire and cable products;
- b) wiring devices;
- c) equipment for use in hazardous locations;
- d) medical electrical equipment and systems;
- e) components\* that will require further evaluation as part of a complete assembly, such as switches, relays, and timers;
- f) equipment connected to line voltage in excess of 46 kV;
- g) manlifts, elevators, climb assists and similar systems (other than their associated control panels); and
- h) any equipment that is not permitted to be field evaluated as directed by an AHJ (such as air-cleaning equipment that intentionally produces ozone).

\*Special inspection/field evaluation only applies to equipment that is capable of being connected to the supply in accordance with the Canadian Electrical Code, Part I.



## SPE-3000 Field Evaluation/Special Inspection

### Inspection label



### Scope of Accreditation:

The performing of field evaluations/special inspections and the application of approval labels on medical-electrical equipment (MEE) and medical-electrical systems (MES)<sup>+</sup> in accordance with all of the following:

As identified in the scope of CSA SPE-3000 Model Code for the field evaluation/special inspection of medical electrical equipment and systems, but excluding the following product categories at this time:

- a) cosmetic and hygiene equipment;
- b) X-ray equipment and systems;
- c) MRIs and CT scans; and
- d) laser equipment.

<sup>+</sup>The definitions for Medical electrical equipment (MEE) and Medical electrical system (MES) can be found in SPE-3000-15 Section 3 Definitions.